

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D0301999	(X3) Date Survey Completed 01/30/2023
Name of Provider or Supplier Drs Dabbs & Hyland P C	Street Address, City, State 1513 Pediatric Drive, Jasper, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on a review of the CASPER reports and American Proficiency Institute (API) proficiency testing records, the laboratory failed to successfully participate in proficiency testing (PT) for White Blood Cell (WBC) Differential for two of three events, Event #1 and #3, 2022. These failures result in an initial unsuccessful PT participation. Refer to D2130.</p>
D2130	<p>HEMATOLOGY CFR(s): 493.851(f)</p>

Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a review of the CASPER reports and American Proficiency Institute (API) proficiency testing records, the laboratory failed to satisfactorily perform for two of three events (Event #1 and #3, 2022) the White Blood Cell (WBC) Differential. The findings include: 1. A review of the CASPER #155 revealed the laboratory scored the following: a) Event #1, 2022, WBC Differential 67 % b) Event #3, 2022, WBC Differential 73 % 2. APT PT records confirmed the above noted findings.