

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D0302153	(X3) Date Survey Completed 04/22/2025
Name of Provider or Supplier Winfield Internal Medicine Pc	Street Address, City, State 644 Tahoe Road, Winfield, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5221	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d)</p> <p>All proficiency testing evaluation and verification activities must be documented.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the Medical Laboratory Evaluation (MLE) Proficiency Testing (PT) records and an interview with Testing Personnel #1, the laboratory failed to document corrective action for scores less than 100 % (percent) for two of six events reviewed in 2023. The findings include: 1. A review of the MLE Chemistry and Non-Chemistry PT records revealed no documentation of corrective action for scores of less than 100% for the following surveys: a) 2023 M2 Chemistry; Urine Sediment 50%, b) 2023 M2 Non-Chemistry; Monocyte% 60%. 2. During an interview on 4/22 /2025, at 11:05 AM, Testing Personnel #1 confirmed the above findings.</p>