

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D0303287	(X3) Date Survey Completed 04/24/2024
Name of Provider or Supplier Urology Specialists, Pc	Street Address, City, State 4704 Whitesburg Drive Suite 100, Huntsville, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on a review of the American Proficiency Institute (API) proficiency test records, corrective action log, and an interview with Testing Personnel #1, the Laboratory failed to submit proficiency test results prior to the event deadline. This was noted for one of six events reviewed in 2023. The findings include: 1. A review of the API proficiency records revealed the laboratory scored zero percent due to "Failure to participate" on the 2023 Hematology 3rd Event. 2. A further review of the API corrective action log revealed "Failure to submit API Heme 3rd event results on due date. Performed blind samples... scored 100%." 3. During an interview on 4/25 /2024, at 10:39 AM, Testing Personnel #1 confirmed the above findings.</p>