

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D0304727	(X3) Date Survey Completed 01/08/2026
Name of Provider or Supplier Affiliated Dermatology	Street Address, City, State 4300 West Main St, Suite 102, Dothan, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5429	<p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(1)</p> <p>(a)(1) Maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the Cryostat maintenance records, the laboratory's policy and procedure manual, and an interview with the MOHS Tech 1 (MT1), the laboratory failed to document Cryostat daily and weekly maintenance as per policy and procedures. This was noted from the date of the last survey (12-13-2023) to the date of the current survey (01-08-2026). The findings include: 1. A review of the Cryostat maintenance records revealed missing documentation of the required nightly and weekly maintenance of the cryostat components for approximately 24 months. 2. A review of the laboratory's policy and procedure manual revealed the following Cryostat activity requirements on page 7. A) "Defrost of machine is done Nightly." B) "The flywheel and moving components on the cryostat are oiled..., every Week." 3. MT1 confirmed these findings during the exit conference on 01-09-2026 at 12 PM.</p>