

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D0668136	(X3) Date Survey Completed 07/26/2023
Name of Provider or Supplier Uahsf Diagnostic Virology Laboratory	Street Address, City, State 1600 7th Ave S Suite 771, Birmingham, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2061	<p>VIROLOGY CFR(s): 493.831(c)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the College of American Pathologists (CAP) proficiency testing (PT) records and an interview with the two General Supervisors, the laboratory failed to ensure PT results were submitted within the time frame specified by the proficiency testing provider. This was noted for one of eighteen 2023 survey events reviewed. The findings include: 1. A review of the CAP PT records revealed no scores and no assessment by CAP for the ID1-A (Nucleic Acid Amplification-Virus) survey due to "failure to submit". 2. A review of the CAP documents shipped with the survey revealed the ID-A survey was received on 4/11/2023, and the submission deadline was 5/2/2023. Records with the survey revealed the survey was performed on 5/2/2023. 3. During an interview on 7/26/2023 at 3:00 PM, General Supervisor (GS) #1 and #2 confirmed these findings. GS #1 stated GS #2 entered the results on the CAP website on 5/2/2023, however she failed to click the "Submit" button because it was her first time performing this task. SURVEYOR ID #32558 Licensure and Certification Surveyor</p>