Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	01D0668755	03/14/2024
Name of Provider or Supplier	**Street Address, City, State**	
Charles W Nolen Jr Md Pc	1705 1st Way Nw, Fayette, AL	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on a review of the Hematology Quality Control (QC) records, a review of the patient log, and an interview with Testing Personnel #4, the laboratory used expired QC reagent on the Beckman Coulter AcT Diff 2 Hematology analyzer. The surveyor noted the laboratory utilized expired QC for two days of patient testing in 2022, 17 patients were affected. The findings include: 1. A review of the Hematology QC records revealed QC was performed on 7/19/2022 and 7/20/2022; all three levels of QC expired 7/18/2022. 2. A review of the patient log revealed 8 patient CBC (Complete Blood Count) tests were performed on 7/19/2022 and 9 patient CBC tests were performed on 7/20/2022. 3. During an interview on 3/14/24, at 12:05 PM, Testing Personnel #4 confirmed the above findings.