

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D0924082	(X3) Date Survey Completed 06/24/2025
Name of Provider or Supplier Childrens Medical Group	Street Address, City, State 610 Providence Park Drive Suite 201, Mobile, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6054	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>(b)(9) Thereafter, evaluations must be performed at least annually</p> <p>This STANDARD is not met as evidenced by: Based on a review of personnel evaluation records, the American Association of Bioanalysts-Medical Laboratory Evaluation (AAB-MLE) Proficiency Testing (PT) records and an interview with the Technical Consultant (TC), the TC failed to assess and document the annual competency for one of the four Testing Personnel (TP) responsible for moderate complexity testing from the date of the last survey, 03-02-2023 to the date of the current survey, 06-24-2025. The findings include: 1. A review of personnel records for TP listed on the CMS-209 Form (Laboratory Personnel Report) revealed the TC failed to perform and document the annual competency assessment for TP4 from 2023-2024. 2. A review of the AAB-MLE PT records revealed TP4 analyzed Non-Chemistry and Chemistry PT samples from the 2023-2024 M1-M3 events. 2. During the exit conference on 06-24-2025 at 1:15 PM, the TC confirmed the above findings.</p>