Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	01D0924324	07/17/2025
Name of Provider or Supplier	**Street Address, City, State**	
Alabama Dermatology And Rejuvenation Center, Pc	1606 Hwy 78 West, Jasper, AL	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5429	MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(1) (a)(1) Maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer. This STANDARD is not met as evidenced by: Based on a review of the Hematoxylin and Eosin (H&E) Histopathology stain maintenance records, the H&E Quality Control and Maintenance procedure, and an interview with the Histopathology Supervisor, the Laboratory failed to perform and document daily and weekly maintenance for the H&E Histopathology stains per the laboratory's procedure. This was noted for 67 days of 2 years reviewed from 2023 through 2025. The findings include: 1. A review of the Hematoxylin and Eosin (H&E) Histopathology stain maintenance records revealed no documentation of daily and weekly maintenance for the following dates: a) 2025: 3/18 and 3/19, b) 2024: 1/16 and 1/17, c) 2023: 10/24, 10/25, November, and December. 2. A review of the H&E Quality Control Maintenance procedure revealed, "Change first alcohol (fixative) and Bluing daily. Change Hematoxylin q 2 weeks... Change Eosin q 2 weeks...." 3. During an interview on 7/17/2025, at 11:11 AM, the Histopathology Supervisor confirmed the above findings.