

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D0997182	(X3) Date Survey Completed 10/06/2026
Name of Provider or Supplier Alabama Dermatology Associates Pc	Street Address, City, State 4410 Watermelon Road, Northport, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on a review of the of bi-annual accuracy verification documentation for KOH (potassium hydroxide), and an interview with the Laboratory Director (LD), the laboratory failed to perform accuracy verification for KOH, a moderate complexity test. This was noted for three of three events reviewed in 2025 and 2026. The findings include: 1. A review of the bi-annual accuracy verification documentation for KOH revealed the laboratory did not perform bi-annual accuracy verification for all testing personnel in the first and second half of 2025. 2. A further review of the bi-annual accuracy verification documentation for KOH revealed the laboratory did not perform bi-annual accuracy verification for the physicians performing KOH for the first half of 2026. 3. The above findings were confirmed by the LD during an interview on 10/6 /26 at 11:56am.