

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  01D0997766	<b>(X3) Date Survey Completed</b>  03/17/2025
<b>Name of Provider or Supplier</b>  Saadat Ansari Md Llc	<b>Street Address, City, State</b>  4810 Whitesport Circle, Sw, Suite 120, Huntsville, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the national database and verified with the laboratory's proficiency testing provider, American Proficiency Institute (API). The laboratory was found to be out of compliance with CONDITION LEVEL DEFICIENCIES, as follows: D2016 - 42 C.F.R. 493.803 Condition: Successful participation [proficiency testing] D6000 - 42 C.F.R. 493.1403 Condition: Laboratories performing moderate complexity testing; laboratory director .
<b>D2016</b>	<p><b>SUCCESSFUL PARTICIPATION</b> CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by:</p>

Based on off-site proficiency testing (PT) desk reviews of the CASPER Reports 0153D and 0155D (Individual Laboratory Profiles from the Centers of Medicare and Medicaid Services [CMS]), and PT evaluation reports from the American Proficiency Institute (API), the laboratory failed to successfully participate (achieve scores of 80% or greater) in proficiency testing for the specialty of Chemistry and the following analytes: Alanine Transaminase (ALT), Total Cholesterol, Creatinine, and Blood Urea Nitrogen (BUN). The laboratory failed two out of three PT events in 2024-2025, resulting in initial unsuccessful proficiency testing performance. Refer to D2088 and D2096. .

**D2088**

**ROUTINE CHEMISTRY**  
CFR(s): 493.841(b)

(b) Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.

This STANDARD is not met as evidenced by:  
Based on off-site proficiency testing (PT) desk reviews of the CASPER Reports 0153D and 0155D (Individual Laboratory Profiles from the Centers of Medicare and Medicaid Services [CMS]), and PT evaluation reports from the American Proficiency Institute (API), the laboratory failed to successfully participate (achieve an overall score of 80% or greater) in proficiency testing for the specialty of Chemistry. The laboratory failed two out of three PT events in 2024-2025, resulting in initial unsuccessful proficiency testing performance. The findings include: 1. A review of the CASPER Reports revealed the laboratory failed to achieve an overall score of 80% or greater in the specialty of Chemistry for the following events: A) 2024 Core Chemistry Event #2: 72% B) 2025 Core Chemistry Event #1: 0% 2. A review of the laboratory's proficiency testing evaluation reports from the API website confirmed these findings. .

**D2096**

**ROUTINE CHEMISTRY**  
CFR(s): 493.841(f)

(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:  
Based on off-site proficiency testing (PT) desk reviews of the CASPER Reports 0153D and 0155D (Individual Laboratory Profiles from the Centers of Medicare and Medicaid Services [CMS]), and PT evaluation reports from the American Proficiency Institute (API), the laboratory failed to successfully participate (achieve scores of 80% or greater) in proficiency testing for Chemistry analytes, Alanine Transaminase (ALT), Total Cholesterol, Creatinine, and Blood Urea Nitrogen (BUN). The laboratory failed two out of three PT events in 2024-2025, resulting in initial unsuccessful proficiency testing performance. The findings include: 1. A review of the CASPER Reports revealed the laboratory received failing scores, as follows: A) 2024 Core Chemistry Event #2 a) ALT: 40% b) Total Cholesterol: 0% c) Creatinine: 40% d) BUN: 40% B) 2025 Core Chemistry Event #1 a) ALT: 0% b) Total

	<p>Cholesterol: 0% c) Creatinine: 0% d) BUN: 0% 2. A review of the laboratory's proficiency testing evaluation reports from the API website confirmed these findings.</p> <p>.</p>
<b>D6000</b>	<p><b>MODERATE COMPLEXITY LABORATORY DIRECTOR</b> CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on off-site proficiency testing (PT) desk reviews of the CASPER Reports 0153D and 0155D (Individual Laboratory Profiles from the Centers of Medicare and Medicaid Services [CMS]), and proficiency testing evaluation reports from the American Proficiency Institute (API), the laboratory director failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D6016. .</p>
<b>D6016</b>	<p><b>LABORATORY DIRECTOR RESPONSIBILITIES</b> CFR(s): 493.1407(e)(4)(i)</p> <p>(e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on off-site proficiency testing (PT) desk reviews of the CASPER Reports 0153D and 0155D (Individual Laboratory Profiles from the Centers of Medicare and Medicaid Services [CMS]), and PT evaluation reports from the American Proficiency Institute (API), the laboratory director failed to ensure the laboratory had successful participation in an HHS approved proficiency testing program for two out of three 2024-2025 API PT events. Refer to D2088 and D2096. SURVEYOR ID #32558 Licensure and Certification Surveyor</p>