

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D1005790	(X3) Date Survey Completed 10/25/2023
Name of Provider or Supplier Huntsville Care Plus, Llc	Street Address, City, State 11100 South Memorial Parkway, Huntsville, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the American Proficiency Institute (API) Proficiency Testing (PT) records and an interview with the Laboratory Supervisor, the laboratory failed to retain the PT review page documentation for five of twenty events reviewed in 2021 through 2022. The findings include: 1. A review of the API PT records revealed no documentation of a review from the Laboratory Director, or designee, for the following surveys: a) 2021 Hematology 3rd Event. b) 2022 Chemistry Miscellaneous 2nd Event. c) 2022 Immunology 1st Event. d) 2022 Immunology 2nd Event. e) 2022 Hematology 1st Event. 2. During an interview on October 25, 2023, at 10:01 AM, the Laboratory Supervisor confirmed the above findings.</p>