

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D1069264	(X3) Date Survey Completed 10/09/2025
Name of Provider or Supplier Whitfield Regional Hospital, Cancer Care Center	Street Address, City, State 105 Us Hwy 80 East, Demopolis, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2121	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>(a) Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the American Proficiency Institute (API) Proficiency Testing (PT) records and an interview with the Technical Consultant (TC) and Testing Personnel 1 (TP1), the laboratory received a failing score for the analyte Hematocrit (HCT) on one of five 2024-2025 PT events. The findings include: 1. A review of the API PT records revealed the laboratory received unsatisfactory scores (60%) for HCT on 2025 Hematology/Coagulation Second Event. 2. During the exit conference on 10-09-2025 at 1:45 PM, TC and TP1 confirmed the above findings.</p>
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on reviews of the API (American Proficiency Institute) Proficiency Testing (PT) records and an interview with the Technical Consultant (TC) and Testing Personnel 1 (TP1), the laboratory failed to ensure PT results were submitted before</p>

the due date specified by API. This was noted for one of the three 2024 Hematology events. The findings include: 1. A review of the API PT records revealed a score of 0 percent due to "Failure to participate" on 2024 Hematology First Event. 2. The TC and TP1 confirmed the above findings during the exit conference on 10-09-2025 at 1:45 PM