Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	01D2017717	09/09/2021
Name of Provider or Supplier	**Street Address, City, State**	
Chamnong Family Medicine	8300 Crossland Loop, Montgomery, AL	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	The surveyor determined this laboratory is in substantial compliance with the requirements of the Clinical Laboratory Improvement Amendments of 1988 (CLIA'88).
D3000	FACILITY ADMINISTRATION CFR(s): 493.1100 Each laboratory that performs nonwaived testing must meet the applicable requirements under 493.1101 through 493.1105, unless HHS approves a procedure that provides equivalent quality testing as specified in Appendix C of the State Operations Manual (CMS Pub. 7). (a) Reporting of SARS-CoV-2 test results During the Public Health Emergency, as defined in 400.200 of this chapter, each laboratory that performs a test that is intended to detect SARS-CoV-2 or to diagnose a possible case of COVID-19 (hereinafter referred to as a "SARS-CoV-2 test") must report SARS-CoV-2 test results to the Secretary in such form and manner, and at such timing and frequency, as the Secretary may prescribe. This CONDITION is not met as evidenced by: Based on record review and interview, the laboratory failed to report SARS-CoV-2 (Severe Acute Respiratory Syndrome Coronovirus 2) test results for positive and negative Coronovirus Disease 2019 (COVID-19) performed on the BinaxNOW COVID-19 Ag Card test kit from May 24, 2021 to September 8, 2021. Findings include: 1. A review of SARS-CoV-2 test results revealed patients were tested for SARS-CoV-2 using the BinaxNOW COVID-19 Ag Card test kit, starting May 24, 2021 through September 8, 2021. A total of 496 test were performed (Positive 88 and Negative 408) during this time period. 2. During an interview on September 09, 2021 at 09:30 AM, the General Supervisor confirmed positive and negative results for SARS-CoV-2 test were not reported to Alabama Department of Public Health.