

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D2044942	(X3) Date Survey Completed 08/26/2025
Name of Provider or Supplier Purohit Pediatric Clinic Roanoke	Street Address, City, State 4441 Highway 431, Suite 1, Roanoke, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A desk review, follow-up survey was conducted on 09-05-2025. Based on a review of an acceptable plan of correction and supporting documentation, submitted by the laboratory and received on 09-10-2025, the surveyor determined the laboratory has achieved substantial compliance with the requirements of the Clinical Laboratory Improvement Amendments of 1988 (CLIA '88).
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on reviews of the API (American Proficiency Institute) Proficiency Testing (PT) records and an interview with Testing Personnel 3 (TP3), the laboratory failed to ensure PT results were submitted before the due date specified by API. This was noted for one of the three 2023 Hematology events. The findings include: 1. A review of the API PT records revealed a score of 0 percent due to "Failure to participate" on 2023 Hematology 3rd Event. 2. TP3 confirmed the above findings during the exit conference on 08-26-2025 at 2:46 PM</p>