

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D2045609	(X3) Date Survey Completed 02/15/2023
Name of Provider or Supplier Southeastern Skin Cancer & Dermatology	Street Address, City, State 104j E Briscoe Way, Madison, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2043	<p>MYCOLOGY CFR(s): 493.827(c)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the American Proficiency Institute (API) Proficiency Testing (PT) records and an interview with Testing Personnel #8, the laboratory failed to submit Mycology Potassium Hydroxide (KOH) PT results within the timeframe specified by the proficiency testing program. This was noted for one of seven events reviewed in 2021 to 2023. The findings include: 1. A review of the API PT records revealed "Failure to participate" for 2022 Mycology 2nd Event resulting in a score of 0% for KOH. The due date for Mycology KOH 2nd event was 7/6/2022 and the lab submitted the results on 7/7/2022. 2. During an interview on February 15, 2023, at 9:41 AM, Testing Personnel #8 confirmed Testing Personnel #5 was unaware of the due date. The lab performed a self evaluation and was successful.</p>