Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 01D2073260	**(X3) Date Survey Completed** 05/13/2021
Name of Provider or Supplier Pain & Rehabilitation Consultants	**Street Address, City, State** 101 Red Hill Way, Madison, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D5215	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2) The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results). This STANDARD is not met as evidenced by: Based on reviews of the proficiency testing (PT) records, quality assurance documentation, patient records, and an interview with the Technical Supervisor, the surveyor determined the laboratory failed to submit results for the API (American Proficiency Institute) 2019-Event 2 Chemistry survey before the cutoff date. This was noted on one out of seven 2018-2021 PT survey events reviewed. The findings include: 1. A review of the instructions for the API 2019-Event 2 Chemistry survey (for qualitative Urine Drug Screens) revealed the PT results should have been submitted on-line on or before 11/1/2019. However a review of the Indiko instrument printouts revealed specimens were run on 11/13/2019; the survey received a score of 0% due to failure to participate. 2. During an interview on 5/13/2021 at 12:15 to 12:25 PM, the Technical Supervisor stated the laboratory may have been out of reagents in October thru early November 2019. The surveyor then asked if there was documentation of the problem, and if patient testing was performed during that period; a review of the quality assurance records revealed no documentation of any problems. The Technical Supervisor confirmed the laboratory performed patient testing on 10/21/2019 and 10/28/2019. SURVEYOR ID #32558 Licensure and Certification Surveyor