

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D2073260	(X3) Date Survey Completed 12/06/2022
Name of Provider or Supplier Pain & Rehabilitation Consultants	Street Address, City, State 101 Red Hill Way, Madison, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on reviews of the API (American Proficiency Institute) proficiency testing (PT) records, and an interview with the Testing Personnel, the surveyor determined the laboratory failed to submit results for the API 2022-Event #2 Chemistry survey before the cutoff date. This was noted on one of four 2021-2022 PT survey events reviewed. This is a repeat deficiency. The findings include: 1. A review of the scores for the API 2022-Event #2 Chemistry survey (for qualitative Urine Drug Screens) revealed the laboratory received a score of 0% due to failure to participate. 2. A review of the instrument printouts for the API 2022-Event #2 Chemistry survey revealed testing was performed on 10/12/2022. The API instruction sheet specified results were due on 11/2/2022 at 11:59 PM ET (Eastern Time). The API records included the "Preview Result Form" with the electronically-entered results, and a printed comment on page one, "THESE TEST RESULTS HAVE NOT BEEN SUBMITTED." 3. During an interview on 12/6/2022 at 12:35 PM, the Testing Personnel confirmed the above findings, and explained there was a miscommunication concerning a text the Technical Consultant sent on 11/1/2022, as follows, "Send in the api b4 11 we'd. My computer is down". The Testing Personnel stated she thought the Technical Consultant meant 11 November, not 11:00 PM on Wednesday. SURVEYOR ID #32558 Licensure and Certification Surveyor</p>