

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D2076788	(X3) Date Survey Completed 10/16/2024
Name of Provider or Supplier Purohit Pediatric Clinic Moody	Street Address, City, State 2468 Moody Pkwy, Moody, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on the tour of the laboratory and an interview with Testing Personnel (TP) #2, the laboratory utilized expired material prior to patient testing. This was noted for 45 days of 2 months reviewed for the Siemens Multistix Reagent Strips, 33 patient urinalysis were affected. The findings include: 1. The laboratory tour with TP#2 revealed the laboratory utilized expired Siemens Multistix Reagent Strips for patient Urinalysis testing. The manufacturer's label indicated Lot# 303016 had an expiration date of 08-31-2024. 2. A further review of the Siemens Clinitek analyzer's memory revealed 33 patients were tested after 08-31-2024 to the date of the survey (10-15-2024). 3. During the exit interview on 10-15-2024 at 1:50 PM, TP#2 confirmed the above findings.</p>