

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  01D2139237	<b>(X3) Date Survey Completed</b>  03/03/2026
<b>Name of Provider or Supplier</b>  Laboratory Corporation Of America Holdings	<b>Street Address, City, State</b>  800 St Vincent'S Drive Suite 500, Birmingham, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D2047</b>	<p><b>PARASITOLOGY</b> CFR(s): 493.829(a)</p> <p>(a) Failure to attain an overall testing event score of at least 80 percent is unsatisfactory performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of the American Proficiency Institute (API) Proficiency Testing (PT) records and an interview with the Laboratory Supervisor (LS), the laboratory failed to achieve successful PT evaluation score for the microscopy vaginal wet preparation event. The surveyor noted one of the three PT events in 2024 had an unsuccessful performance. The findings include: 1. A review of the API PT records revealed the 2024 Hematology and Coagulation First event for the microscopy vaginal wet preparation was zero percent. 2. The LS confirmed the above findings on 03-03-2026 at 1:31 PM during the exit conference.</p>
<b>D6053</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(9)</p> <p>(b)(9) Evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on a review of personnel records and an interview with the Laboratory Supervisor (LS), the Technical Consultant (TC) failed to assess competency of Testing Personnel (TP) at least semi-annually during the first year of patient testing. The surveyor noted one of the five TP who performed moderate complexity testing had no documentation of the required semi-annual assessment in 2024. The findings</p>

include: 1. A review of personnel records revealed TP3 listed on the CMS 209 (Laboratory Personnel Report) had no documentation of the 2024 semi-annual competency assessment available for review during the survey. 2. The LS confirmed the above findings on 03-03-2026 at 1:31 PM during the exit conference.