Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 01D2208265	**(X3) Date Survey Completed** 12/08/2025
Name of Provider or Supplier Huntsville Mens Clinic	**Street Address, City, State** 250 Chateau Drive Sw, Suite 150, Huntsville, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	The following deficiencies are a result of a desk review of proficiency testing scores obtained from the national database and verified with the laboratory's proficiency testing provider, American Proficiency Institute (API). The laboratory was found to be out of compliance with CONDITION LEVEL DEFICIENCIES, as follows: D2016 - 42 C.F.R. 493.803 Condition: Successful participation [proficiency testing] D6000 - 42 C.F.R. 493.1403 Condition: Laboratories performing moderate complexity testing; laboratory director.
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on proficiency testing (PT) desk reviews of the CASPER Reports 0153D and

	0155D (Individual Laboratory Profiles from the Centers of Medicare and Medicaid Services [CMS]), and PT evaluation reports from the American Proficiency Institute (API), the laboratory failed to successfully participate (achieve scores of 80% or greater) in proficiency testing for the specialty Chemistry and PSA (Prostate Specific Antigen), an analyte in the specialty of Chemistry. The laboratory failed two PT events in 2025, resulting in initial unsuccessful proficiency testing performance. Refer to D2096 and D2097.
D2096	ROUTINE CHEMISTRY CFR(s): 493.841(f) (f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on proficiency testing (PT) desk reviews of the CASPER Reports 0153D and 0155D (Individual Laboratory Profiles from the Centers of Medicare and Medicaid Services [CMS]), and PT evaluation reports from the American Proficiency Institute (API), the laboratory failed to successfully participate (achieve scores of 80% or greater) in proficiency testing for PSA (Prostate Specific Antigen). The laboratory failed two PT events in 2025, resulting in initial unsuccessful proficiency testing performance. The findings include: 1. A review of the CASPER Reports revealed the laboratory received failing scores for PSA in two API PT events, as follows: A) 2025 Event 1: PSA 20% B) 2025 Event 3: PSA 20% 2. A review of the laboratory's proficiency testing evaluation reports provided by API confirmed these findings.
D2097	ROUTINE CHEMISTRY CFR(s): 493.841(g) (g) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance. This STANDARD is not met as evidenced by: Based on a review of the Centers for Medicare and Medicaid Services (CMS) CASPER Reports and American Proficiency Institute (API) testing evaluations, the laboratory failed to satisfactorily perform for two Chemistry testing events (Event #1 and #3, 2025), for PSA (Prostate Specific Antigen). These failures resulted in an unsuccessful performance. The findings include: 1. A review of the CASPER #155 revealed the laboratory scored 20% for Chemistry Event #1 and #3, 2025. 2.The API proficiency testing evaluations confirmed the above noted score.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493. 1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.

	This CONDITION is not met as evidenced by: Based on proficiency testing (PT) desk reviews of the CASPER Reports 0153D and 0155D (Individual Laboratory Profiles from the Centers of Medicare and Medicaid Services [CMS]), and proficiency testing evaluation reports from the American Proficiency Institute (API), the laboratory director failed to ensure successful participation in an HHS approved proficiency testing program. Refer to D6016.
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) (e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part; This STANDARD is not met as evidenced by: Based on proficiency testing (PT) desk reviews of the CASPER Reports 0153D and 0155D (Individual Laboratory Profiles from the Centers of Medicare and Medicaid Services [CMS]), and PT evaluation reports from the American Proficiency Institute (API), the laboratory director failed to ensure the laboratory had successful participation in an HHS approved proficiency testing program for the specialty Chemistry and the analyte PSA (Prostate Specific Antigen) in the specialty of Chemistry for two 2025 API PT events. Refer to D2096 and D2097.