

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 01D2275483	(X3) Date Survey Completed 12/19/2025
Name of Provider or Supplier Med 360 Urgent Care Northport Llc	Street Address, City, State 3250 Mcfarland Blvd, Northport, AL	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the Hematology Quality Control (QC) records, and an interview with the Technical Consultant, the laboratory used expired QC reagent on the Sysmex XP-300 Hematology analyzer prior to patient testing. The surveyor noted the laboratory utilized expired QC for 6 of 31 days of patient testing in December 2023. The findings include: 1. A review of the Hematology QC records revealed the QC was performed on 12/14/2023 through 12/19/2023; all three levels of QC expired 12/13 /2023. 2. During an interview on 12/19/2025, at 2:12 PM, the Technical Consultant confirmed the above findings.</p>