Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 02D1031794	**(X3) Date Survey Completed** 12/06/2018
Name of Provider or Supplier Alaska Regional Hospital Poc Testing	**Street Address, City, State** 2801 Debarr Road, Anchorage, AK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D5775	COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c) (a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. (c) The laboratory must document all test result comparison activities. This STANDARD is not met as evidenced by: Based on review of test records and interview with the technical consultant, the laboratory did not perform, document, and evaluate the comparison of test results for tests that are performed on multiple instruments at least twice a year. Findings: 1. The laboratory performs pH, pO2, pCO2, sodium, potassium, ionized calcium, glucose, lactate, and hematocrit on both the IL Gem Premier 3500 and Abbott i-STAT CG4+ and CCG8+ cartridges. 2. The laboratory performs Activated Clotting Times (ACT) on both the Medtronic ACT and Abbott i-STAT ACT cartridges. 3. The laboratory has 18 Abbott i-STATs, 1 IL Gem Premier, 1 Metronic ACT, and performs approximately 25,000 chemistry tests and 1,000 ACTs annually. 4. Missing were comparisons of test results for these analytes. 5. The technical consultant confirmed these findings on 12/6/2018 at 13:30.