

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 02D1090167	(X3) Date Survey Completed 12/03/2018
Name of Provider or Supplier Lifemed Alaska Llc	Street Address, City, State 6320 S Airpark Place, Anchorage, AK	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6065	<p>TESTING PERSONNEL QUALIFICATIONS CFR(s): 493.1423(b)(1)(2)(3)(4)(i)</p> <p>(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located or have earned a doctoral, master's, or bachelor's degree in a chemical, physical, biological or clinical laboratory science, or medical technology from an accredited institution; or (b)(2) Have earned an associate degree in a chemical, physical or biological science or medical laboratory technology from an accredited institution; or (b)(3) Be a high school graduate or equivalent and have successfully completed an official military medical laboratory procedures course of at least 50 weeks duration and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); or (b)(4)(i) Have earned a high school diploma or equivalent; and</p> <p>This STANDARD is not met as evidenced by: Based on review of testing personnel records and technical consultant interview, the laboratory did not have copies of diplomas, transcripts or primary source verification (PSV) for the testing personnel. Findings: 1. There were 39 new testing persons hired since the last survey in 2016. 35 of them did not have copies of their diploma, transcript, or primary source verification. 2. The laboratory performs approximately 1300 tests annually. 3. The technical consultant and laboratory director confirmed these findings on 12/3/18 at 12:00 pm.</p>