

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D0530945	(X3) Date Survey Completed 05/22/2023
Name of Provider or Supplier All Dermatology	Street Address, City, State 6320-A W Union Hills Dr Ste 210, Glendale, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation of reagents and interview with the facility personnel, the laboratory used the reagent, Potassium Hydroxide (KOH), past the expiration date. Findings include: 1. The laboratory performs KOH testing on patient specimens under the subspecialty of Mycology, with an approximate annual test volume of 100. 2. During the survey conducted on May 22, 2023, direct inspection of the Potassium Hydroxide reagent used for patient testing (lot #8506-01) indicated an expiration date of 7/31/2020. 3. The number of patients tested using the expired reagent could not be determined at the time of the survey. 4. At approximately 10:25am on 5/22/2023, the facility personnel interviewed stated that the expired reagent indicated above was in use at the time of the survey.</p>