

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  03D0531824	<b>(X3) Date Survey Completed</b>  09/11/2024
<b>Name of Provider or Supplier</b>  Benson Hospital	<b>Street Address, City, State</b>  450 S Ocotillo, Benson, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5211</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of Proficiency Testing (PT) records from 2023 and interview with the technical supervisor (TS-1), the laboratory failed to provide a documented review of the PT results for testing events in the specialty of Chemistry. Findings include: 1. The laboratory performs patient testing in the specialty of Chemistry with an approximate annual test volume of 15,673. 2. The laboratory participates in PT testing through American Proficiency Institute (API) for all reported analytes. It participates in three events annually for all regulated analytes. 3. No evidence, either by written comment or signature, was presented during the survey conducted on 9/11/24 to indicate the laboratory director or the designated technical supervisor reviewed the PT results for the first and second Chemistry Core testing events of 2023. 4. The TS-1 interviewed on 9/11/24 at 10:10 AM confirmed that the review of the PT results indicated above were not documented.</p>