

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D0912335	(X3) Date Survey Completed 07/22/2026
Name of Provider or Supplier Associated Dermatologists Pc	Street Address, City, State 1055 N La Canada, Ste 125, Green Valley, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5291	GENERAL LABORATORY SYSTEMS QUALITY ASSESSMENT CFR(s): 493.1239(a) The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and, when indicated, correct problems identified in the general laboratory systems requirements specified at 493.1231 through 493.1236. This STANDARD is not met as evidenced by: Based on a review of the policy and procedure manual and interview with the facility personnel on 7/22/26 at 12:15 PM, the laboratory failed to establish correct policies and procedures for the accuracy verification process of Mohs specimens and frozen biopsies. Findings include: 1. The laboratory performs patient testing under the subspecialty of Histopathology with a reported annual test volume of 700. The laboratory performs the microscopic interpretation of Mohs specimens and frozen biopsies. 2. A review of the laboratory's "Proficiency Testing/Peer Review Policy" was performed on 7/22/26. The policy states: Twice yearly, one of the Mohs surgery medical assistants will randomly select a minimum of two micrographic surgery cases and one frozen section biopsy from the prior year for peer review." 3. The CLIA regulation, 493.1236(c)(1) states: "At least twice annually (current year), the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. 4. The facility personnel interviewed on 7/22/26 at 12: 15 PM confirmed the laboratory failed to establish correct policies and procedures for the accuracy verification process of Mohs specimens and frozen biopsies.