

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D0955419	(X3) Date Survey Completed 01/12/2024
Name of Provider or Supplier Southwest Skin Specialists	Street Address, City, State 11130 N Tatum Blvd #I-100, Phoenix, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on lack of accuracy verification documentation for Dermatophyte Test Medium (DTM) testing and interview with the Laboratory Director (LD), the laboratory failed to verify the accuracy of testing performed under the subspecialty of Mycology at least twice annually during 2022 and 2023. Findings include: 1. No documentation was presented for review to indicate the laboratory verified the accuracy of DTM testing at least twice annually during 2022 and 2023. 2. The LD interviewed on January 12, 2024 at 11:00 AM confirmed the laboratory failed to verify the accuracy of DTM testing at least twice annually during 2022 and 2023. 3. The laboratory performs approximately 20 DTM tests annually under the subspecialty of Mycology.</p>