

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D0992206	(X3) Date Survey Completed 01/16/2025
Name of Provider or Supplier Arizona Dermatology	Street Address, City, State 6644 E Baseline Rd Ste 102, Mesa, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on lack of accuracy verification documentation for Mohs and interview with the testing personnel, the laboratory failed to verify the accuracy of testing performed under the subspecialty of Histopathology at least twice annually during 2023 and 2024. Findings include: 1. No documentation was presented for review to indicate the laboratory verified the accuracy of Mohs testing at least twice annually during 2023 and 2024. 2. The testing personnel interviewed on 1/16/25 at 11:40 AM confirmed the laboratory failed to verify the accuracy of Mohs testing testing at least twice annually during 2023 and 2024. 3. The laboratory's reported annual test volume under the subspecialty of Histopathology is 1,658.</p>