Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D1074397	(X3) Date Survey Completed 11/19/2024
Name of Provider or Supplier Skin & Cancer Center Of Arizona Pc	**Street Address, City, State** 725 S Dobson Rd, Room 202, Chandler, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on direct observation of the Potassium Hydroxide (KOH) and Chlorazol Black E reagents and interview with the facility personnel, the laboratory used the Potassium Hydroxide, 10% and Chlorazol Black E reagents past the expiration dates. Findings include: 1. The laboratory performs KOH preparations on patient slides under the subspecialties of Mycology and Parasitology with a reported annual test volume of 220. 2. Direct inspection of the Potassium Hydroxide, 10% reagent (lot #2545-03) in use at the time of the survey conducted on 11/19/24 indicated an expiration date of 8 /31/23. 3. Direct inspection of the Chlorazol Black E reagent (lot# 1120) in use at the time of the survey conducted on 11/19/24 indicated an expiration date of 4/30/23. 4. The total number of patients tested using the expired reagents could not be determined at the time of the survey. 5. The facility personnel interviewed on 11/19/24 at 9:10 AM confirmed the expired reagents indicated above were in use at the time of the survey.