

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D1098554	(X3) Date Survey Completed 02/13/2025
Name of Provider or Supplier Integrated Dermatology Of Yuma	Street Address, City, State 2500 S 8th Ave Ste 101, Yuma, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>(b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.</p> <p>This STANDARD is not met as evidenced by: Based on lack of room temperature records from October 6, 2023 through November 8, 2024 and interview with the facility personnel, the laboratory failed to monitor and document the room temperature where dermatopathology reagents and equipment are utilized and stored on each day of patient testing. Findings include: 1. The laboratory processes specimens and interprets dermatopathology slides in conjunction with Mohs surgery, with an approximate annual test volume of 250. 2. The laboratory failed to monitor and document the temperature of the room where dermatopathology reagents and equipment are utilized each day of testing on 26 out of 26 testing dates between October 6, 2023 through November 8, 2024. 3. Mohs records (including patient logs) reviewed during the survey indicate 250 patients were tested during the timeframe indicated above. 4. The facility personnel interviewed on 2/13/25 at 11:25 AM confirmed that the laboratory failed to monitor and document the room temperature of the laboratory from October 6, 2023 through November 8, 2024.</p>