

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D2008875	(X3) Date Survey Completed 04/11/2019
Name of Provider or Supplier Csl Plasma, Inc- Tucson	Street Address, City, State 4775 S Butterfield Dr Ste 175-195, Tucson, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on lack of documentation provided by the laboratory from the proficiency testing (PT) organization and interview with the laboratory personnel, the laboratory failed to evaluate the results of the PT samples for the 3rd event of 2018 for Total Protein testing compared to the results supplied by the PT organization. Findings include: 1. The laboratory received an unsatisfactory score of 0 for the third event of 2018 for Total Protein testing since the PT results were not submitted to the PT organization by the deadline. 2. The laboratory provided a corrective action for the late submission of the PT results. 3. The laboratory provided no assessment of the PT results from the testing of the PT samples compared to the actual results supplied by the PT organization. 4. The laboratory personnel presented a blank document from the PT organization that would normally contain the results utilized for comparison and assessment. The laboratory personnel indicated that the results would not print out. 5. Prior to the survey conducted on 04/11/2019, the PT organization was not contacted by the laboratory for follow-up regarding obtaining the actual results for comparative analysis.</p>