

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D2024481	(X3) Date Survey Completed 03/21/2018
Name of Provider or Supplier Saguaro Dermatology Ahwatukee	Street Address, City, State 4425 E Agave Rd, Ste 148, Phoenix, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of accuracy verification documentation, laboratory policy review and interview with the facility personnel, the laboratory failed to verify the accuracy of dermatopathology testing at least twice annually in 2017. Findings include: 1. The laboratory performs patient testing under the sub-specialty of Histopathology, with an approximate annual test volume of 200. 2. The laboratory presented accuracy checks for Mohs cases that were reported in 2017, but were not reviewed by an outside dermatopathologist until March of 2018. There were no cases sent out for review in 2017. 3. The laboratory policy indicates that 5 cases selected at random are to be reviewed by an outside dermatopathologist biannually. 4. The facility personnel acknowledged that no cases were sent out for review in 2017.</p>