

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  03D2066117	<b>(X3) Date Survey Completed</b>  09/24/2020
<b>Name of Provider or Supplier</b>  Bingham Dermatology Group	<b>Street Address, City, State</b>  2855 E Brown Rd Ste 28, Mesa, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on review of patient logs, direct observation of staining reagents and interview with the facility personnel, the laboratory utilized Xylene staining reagent past the expiration date indicated on the reagent label. Findings include: 1. Direct observation of the Xylene reagent used in the staining process of Mohs surgical slides revealed an expiration date of 10/07/2018. 2. The facility personnel confirmed that the expired reagent indicated above was being used in the slide staining process up to the day of the survey. 3. The number of patients tested using the expired Xylene reagent could not be determined at the time of the survey.</p>