

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  03D2113129	<b>(X3) Date Survey Completed</b>  03/03/2026
<b>Name of Provider or Supplier</b>  Mayo Clinic Primary Care - Phoenix	<b>Street Address, City, State</b>  5701 E Mayo Blvd, Phoenix, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An onsite validation survey was conducted on 03/03/2026, and standard level deficiencies were cited.
<b>D5775</b>	<p><b>COMPARISON OF TEST RESULTS</b> CFR(s): 493.1281(a)(c)</p> <p>(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation, a review of the laboratory's policies and procedures, and an interview with Technical Consultant 1 (TC1), the laboratory failed to perform and document twice a year comparison for two of two years (2024-2025). a. On 03/03/2026 at 1:08 PM, in the laboratory, the surveyor observed two CoaguChek XS Pro Systems with serial numbers 9812 and 3775. b. The laboratory was asked to provide documentation and evidence of twice-a-year instrument comparisons. No documentation was provided. c. An interview with TC1 on 03/03/2026 at 1:46 PM in the conference room confirmed that no comparison was made between the two instruments.</p>