

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D2128012	(X3) Date Survey Completed 06/07/2018
Name of Provider or Supplier Premier Medical Group Laboratory	Street Address, City, State 5005 S Ash Ave Suite A-2, Tempe, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on lack of calibration verification documentation for the Beckman AU680 chemistry analyzer and interview with the facility personnel, the laboratory failed to perform and document calibration verification procedures as required. Findings include: 1. The electrolytes included in testing of chemistry panels only include a two point calibration 2. No documentation was presented for review for October or November of 2017 to indicate the laboratory performed a calibration verification at</p>

least once every six months for the electrolytes tested under routine chemistry, including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results. The laboratory provided documentation of a calibration verification performed in April 2017 as part of the performance verification study and in May 2018, but no other documentation was presented for review. 3. The facility personnel confirmed that the laboratory did not perform a calibration verification every six months as required.