

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 03D2303946	(X3) Date Survey Completed 02/26/2026
Name of Provider or Supplier Peak Dermatology	Street Address, City, State 16838 E Palisades Blvd Ste 105, Fountain Hills, AZ	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation of Histopathology stain reagents and interview with the facility personnel, the laboratory used the stain reagent, Scott Tap Water Substitute, past the expiration date of 4/18/2025. Findings include 1. The laboratory processes dermatopathology tissues and performs the microscopic interpretation of slides in the subspecialty of Histopathology with a reported annual test volume of 150. 2. During the survey conducted on 2/26/26, direct inspection of the Scott Tap Water Substitute, lot # 2409905, indicated an expiration date of 4/18/25. 3. The laboratory used the expired reagent to process patients' slides throughout the timeframe of 4/19/25 through 2/26/26. 90 patients were tested during this timeframe. 4. The facility personnel interviewed on 2/26/26 at 9:40 AM confirmed the expired reagent indicated above was used for patient testing past the manufacturer's expiration date and was in use at the time of the survey.</p>