

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  04D0465998	<b>(X3) Date Survey Completed</b>  07/29/2025
<b>Name of Provider or Supplier</b>  Bh Arkadelphia Medical Clinic	<b>Street Address, City, State</b>  2913 Cypress Road, Suite 100, Arkadelphia, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 7/29/25 at 10:55 am, thirty BD ESwab Collection &amp; Transport System for Aerobic, Anaerobic &amp; Fastidious Bacteria Lot # 1440069000 expiration date 2025-07-12 were observed in a laboratory cabinet, available for use beyond the expiration date. B) In an interview on 7/29/25 at 11:00 am the laboratory staff member ( #5 on the form CMS 209) confirmed that the items, identified above, had exceeded their expiration dates and were available for use in the laboratory.</p>