

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  04D0466223	<b>(X3) Date Survey Completed</b>  01/07/2025
<b>Name of Provider or Supplier</b>  Clinton Medical Clinic	<b>Street Address, City, State</b>  2526 Hwy 65 South, Suite 203, Clinton, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 1/07/25 at 11:20 am, one (of one) opened container of SCS1000 Hematology Calibrator (lot 4107, REF 217685, Sysmex, expiration date 5/19/24) was observed in the laboratory, available for use beyond the expiration date. B) In an interview on 1/7/25 at 11:24 am the lab director, confirmed that the item, identified above, had exceeded the expiration date and was available in the laboratory.</p>