

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D0467525	(X3) Date Survey Completed 01/06/2023
Name of Provider or Supplier Central Arkansas Pediatric Clinic	Street Address, City, State 2301 Springhill Road, Suite 200, Benton, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Through review of package insert, observation, and interview it was determined that laboratory supplies were available for use when they have exceeded their expiration date. Finding follows: A) Review of the eSwab package insert states do not use after expiration date, which is printed on the outer box and on each individual sterile collection unit and the specimen transport tube label. B) On 01/06/2023 at 1209, 10 of 10 eSwab transport system tubes for the collection and preservation of aerobic, anaerobic, and fastidious bacteria were found in a drawer. The expiration date on all eSwab 2022-09-30 (lot number 211631000). C) In an interview on 1/06/2023 at 12:25 laboratory staff member (number 21 on form CMS-209) confirmed that eSwab were accessible for use for all testing personnel when they had exceeded the expiration date.</p>