

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D0467710	(X3) Date Survey Completed 03/25/2025
Name of Provider or Supplier Family Medicine Assoc Of Blytheville	Street Address, City, State 1521 N Tenth Suite C, Blytheville, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of the 2024 and 2025 CMS Casper Reports 096D, 0153D, and the American Proficiency Institute (API) proficiency testing results, the laboratory failed to have initial successful participation in proficiency testing for the analyte Thyroid-stimulating Hormone (TSH) and the subspecialty of Endocrinology. Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance as cited at D2107, and D2108.</p>
D2107	ENDOCRINOLOGY

	<p>CFR(s): 493.843(f)</p> <p>(f) Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of the 2024 and 2025 CMS Casper Reports 096D, 0153D and API proficiency testing results, it was determined the laboratory failed to have satisfactory participation in proficiency testing for the analyte TSH. Survey Findings follow: A. A review of the proficiency testing results revealed the laboratory received a score of 60% for the analyte TSH in the third proficiency testing event of 2024. B. A review of the proficiency testing results revealed the laboratory received a score of 20% for the analyte TSH in the first proficiency testing event of 2025.</p>
<p>D2108</p>	<p>ENDOCRINOLOGY CFR(s): 493.843(g)</p> <p>(g) Failure to achieve an overall testing event score of satisfactory performance for two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of the 2024 and 2025 CMS Casper Reports 096D, 0153D and API proficiency testing results, it was determined the laboratory failed to have satisfactory participation in proficiency testing for the subspecialty of Endocrinology. Survey Findings follow: A. A review of the proficiency testing results revealed the laboratory received a score of 60% for the subspecialty of endocrinology in the third proficiency testing event of 2024. B. A review of the proficiency testing results revealed the laboratory received a score of 20% for the subspecialty of endocrinology in the first proficiency testing event of 2025.</p>
<p>D6000</p>	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on review of 2024 and 2025 proficiency testing results, the laboratory director failed to ensure that the proficiency testing samples are tested as required under Subpart H of this part. Refer to D6016.</p>
<p>D6016</p>	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i)</p> <p>(e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part;</p>

This STANDARD is not met as evidenced by:
Based on review of the 2024 proficiency testing events, the laboratory director failed to ensure the laboratory successfully participated in proficiency testing for the analyte TSH, and the overall subspecialty of Endocrinology. Refer to D2107, and D2108.