

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  04D0467710	<b>(X3) Date Survey Completed</b>  02/25/2026
<b>Name of Provider or Supplier</b>  Family Medicine Assoc Of Blytheville	<b>Street Address, City, State</b>  1521 N Tenth Suite C, Blytheville, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 2/25/26 at 2:17 pm, 14 tubes of BD Mircotainer MAP K2E (1.0mg) REF 363706 Lot 4169587, expiration date 12-31-2025 was observed in the laboratory, available for use beyond the expiration date. B) In an interview on 2/25/26 at 2:27 pm the technical consultant confirmed that the item, identified above, had exceeded it's expiration date and was available for use in the laboratory</p>