

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D0642286	(X3) Date Survey Completed 01/15/2025
Name of Provider or Supplier Nea Baptist Clinic - Main Laboratory	Street Address, City, State 4802 E Johnson Ave, Jonesboro, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 1/15/25 at 3:10 pm, one (of one) opened container of Wright Stain, Modified (lot SLCL9514, REF WS128-4L, Sigma-Aldrich, expiration date 8/1/2023) was observed in the laboratory, available for use beyond the expiration date. B) In an interview on 1/15/25 at 3:21 pm the lab director, confirmed that the item, identified above, had exceeded the expiration date and was available in the laboratory.</p>