

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D0697597	(X3) Date Survey Completed 07/22/2026
Name of Provider or Supplier St Bernards Physician Clinics, Inc	Street Address, City, State 21110 Hwy 18 East, Monette, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) (d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based upon observation and interview the laboratory had fifty blood collection tubes that had past the date of expiration available for use. Findings follow: A) During a tour of the laboratory on 7/22/26 at 03:52 p.m., 50 BD Vacutainer Lithium Heparin 3.0 ml. blood collection tubes (lot # 5169938, expiration date 2026-06-30) were observed next to the phlebotomy chair in the laboratory. B) In an interview on 7/22/26 at 04:00 p.m., the laboratory staff member (#2 on the form CMS 209) confirmed the tubes identified above were expired and were availble for use.