

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 04D0896920	<b>(X3) Date Survey Completed</b> 05/07/2026
<b>Name of Provider or Supplier</b> Uams Regional Programs Jonesboro	<b>Street Address, City, State</b> 311 E Matthews Ave, Jonesboro, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 5/7/26 at 1:09pm, 1 of 45 CL Cell Clean Auto cleaing agent, lot: A5049, REF CF-579-595, expiration date 4/29/26, was available for use beyond the expiration date. B) In an interview on 5/7/26 at 1:10pm the technical consultant confirmed that the item, identified above, had exceeded it's expiration date and was available for use in the laboratory.</p>