

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D0906636	(X3) Date Survey Completed 12/07/2023
Name of Provider or Supplier Family Practice Assocs Of Clarksville	Street Address, City, State 23 Professional Park Drive, Clarksville, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Through observation and interview with laboratory staff it was determined that the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 12/07/23 at 2:46 p.m. one (of one) ISE Cleaning Solution / Elecsys Sysclean (Lot: 62235301, ref: 11298500160, expiration date 11/30/23) was observed in the laboratory. C) In an interview on 12/07/23 at 2:46 p.m laboratory staff member (listed as #3 on the CMS 209 form) confirmed that the items, identified above, had exceeded their expiration date and were available for use.</p>