

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D0980285	(X3) Date Survey Completed 11/16/2023
Name of Provider or Supplier Little Rock Children's Clinic P A	Street Address, City, State 27 Corporate Hill Drive, Suite 100, Little Rock, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Through observations made during a tour of the laboratory, it was determined the laboratory had reagents available for use when they had exceeded their expiration date. Survey findings follow: A) During a tour of the laboratory at 12:45 p.m.on 11/16 /23 ten of ten Eosin-Methylene Blue (EMB) microbiology media agar plates, lot # 23423 expiration date 11/14/23, were observed available for use when they had exceeded their expiration date. B) In an interview on 11/16/23 at 12:45 p.m., the laboratory staff member (# 1 on the CMS 209 form) confirmed that the media identified above had exceeded the expiration date and was available for use.</p>