

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D1007429	(X3) Date Survey Completed 06/25/2026
Name of Provider or Supplier Healthy Connections	Street Address, City, State 136 Health Park Drive, Mena, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d) (d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality. This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 6/26/26 at 3:29pm, two (of two) containers of "BD Vacutainer Safety-Lok Blood Collection Set" (BD, lot: 2D2481, REF: 367285, expiration date 2024-04-30) were observed in the laboratory, available for use beyond the expiraton date. B) In an interview on 6/26/26 at 3:31pm the testing personnel # 2 (as identified on the CMS-209 form) confirmed that the items, identified above, had exceeded their expiration date and were available for use in the laboratory but had not been used for collecting patient samples.