

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  04D1039733	<b>(X3) Date Survey Completed</b>  07/17/2025
<b>Name of Provider or Supplier</b>  Unity Health Searcy Medical Center West Clinic	<b>Street Address, City, State</b>  2505 West Beebe Capps Expressway, Searcy, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 7/17/25 at 12:58pm, three (of three) containers of Eprexia Resolve Immersion Oil (lot 113323, REF M3004, expiration date 12/2023) were observed in the laboratory, available for use beyond the expiration date. C) In an interview on 7/17/25 at 12:59pm testing personnel #1 (as listed on the CMS-209 form), confirmed that the items, identified above, had exceeded their expiration dates and were available for use in the laboratory.</p>