

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D1051973	(X3) Date Survey Completed 07/23/2026
Name of Provider or Supplier Nwa Pathology At Wrnc Fs Room	Street Address, City, State 3214 North North Hills Blvd, Fayetteville, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5413	TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b) (b) The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (b)(1) Water quality. (b)(2) Temperature. (b)(3) Humidity. (b)(4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports. This STANDARD is not met as evidenced by: Based upon observation, review of temperature records, lack of documentation, and interview the laboratory failed to monitor the temperature and humidity on each day of operation. . Findings follow: A) During a tour of the laboratory on 7/23/26 at 9: 43am a Dakewe Cryostat with operating temperature and humidity requirements was observed. B) A review of the laboratory's temperature records revealed that no room temperatures or humidities were recorded for the laboratory for 75 of 75 days of operation in 2026. C) The Dakewe Cryostat 6250 User's Manual (Revised Version V5. 1-English Version-Dec.2023) states "Installation Requirements, Room temperature is always between 5C and 35C, Relative humidity is up to 60% without condensation." D) Upon request, the laboratory could not present the temperature and humidity records for 75 of 75 days of operation for 2026. E) In an interview on 7/23/26 at 9: 58am, the laboratory staff member (#2 on the entrance/exit conference sign in sheet) confirmed that temperature and humidity records were not available.