

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 04D1073729	(X3) Date Survey Completed 12/09/2025
Name of Provider or Supplier Ozark Dermatology Clinic	Street Address, City, State 901 Se 22nd St, Bentonville, AR	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with laboratory staff, the laboratory had supplies available for use after their expiration date. Findings follow: A) During a tour of the laboratory on 12/9/25 at 12:41 pm, thirty-five BD 4.5 ml lithium/heparin blood collection tubes lot# 4198313 expiration date 2025-08-31, 12 BD Trace element blood collection tubes lot# 4222034 expiration date 2025-07-31 and one bottle of CoverSeal-X lot# 112865 expiration date 11/28//25 were observed in a the laboratory, available for use beyond the expiraton date. B) In an interview on 12/9/25 at 02:33 pm the laboratory staff member (#3 on the form CMS 209) confirmed that the items, identified above, had exceeded their expiration dates and were available for use in the laboratory.</p>